



ENVIRONMENT
AGENCY

PRODUCT CONFORMITY CERTIFICATE

This is to certify that the

TFID Thermo Flame Ionisation Detector

manufactured by:

Emerson Process Management

*Manufacturing GmbH & Co.OHG
Industrie Straße 1
D-63594 Hasselroth
Germany*

has been assessed by Sira Certification Service
and for the conditions stated on this certificate complies with:

MCERTS Performance Standards for Continuous Emission Monitoring Systems (April 2003)

Certification Ranges :

TOC 0 to 15 mgC/m³
 0 to 30 mgC/m³

Project No: 674/0137
Certificate No: Sira MC 050054/02
Initial Certification: 04 February 2005
This Certificate Issued: 23 August 2011
Renewal Date: 03 February 2010

Technical Director

MCERTS is operated on behalf of the Environment Agency by

Sira Certification Service

12 Acorn Industrial Park, Crayford Road, Crayford
Dartford, Kent, UK, DA1 4AL
Tel: 01322 520500 Fax: 01322 520501

This certificate may only be reproduced in its entirety and without change



ENVIRONMENT
AGENCY

Certified Performance

The instrument was evaluated for use under the following conditions:

Ambient Temperature Range: +5°C to +40°C

Performance values are expressed as a percentage of the certification range of 0 to 15 mgC.m⁻³, except for availability and analysis function, and '✓' Indicates compliance with MCERTS requirements.

Test	Results expressed as % of certification range				Other results	MCERTS specification
	<0.5	<1	<2	<4		
Detection limit	✓				<0.33mg C.m ⁻³	<0.4mg C.m ⁻³
Linearity	✓		✓		0.17mg C.m ⁻³ 1.3%	<0.4mg C.m ⁻³ 5% of range (0-30mgC.m ⁻³)
Response time					10s	<60s
Effect of oxygen					0.53mgC.m ⁻³	<0.8mgC.m ⁻³
Range of response factors:						
Aliphatic hydrocarbons					1.0-1.06	0.90-1.10
Aromatic hydrocarbons					0.91-1.0	0.85-1.1
Dichloromethane					0.86-0.87	0.75-1.15
Aliphatic alcohols					0.75-0.87	0.7-1.0
Esters and ketones					0.74-0.76	0.7-1.0
Cross-sensitivity					<0.59mgC.m ⁻³	<±1.0mgC.m ⁻³
Effect of CO ₂ , CO, SO ₂ , NH ₃ , N ₂ O, HCl, H ₂ O, CH ₄						

Certificate No: Sira MC 050054/02
This Certificate Issued: 23 August 2011

This certificate may only be reproduced in its entirety and without change



**ENVIRONMENT
AGENCY**

Test	Results expressed as % of certification range				Other results	MCERTS specification
	<0.5	<1	<2	<4		
Zero drift (field)		✓				<±2%/week
Span drift (field)			✓			<±4%/week
Temperature responsive zero shift	✓				<0.14 %/°C	<±0.3%/°C
Temperature responsive span shift	✓				<0.22 %/°C	<±0.3%/°C
Availability					98%	>95%
Reproducibility					35	≥30
Maintenance interval					4 weeks	To be reported

Certificate No: Sira MC 050054/02
 This Certificate Issued: 23 August 2011

This certificate may only be reproduced in its entirety and without change



ENVIRONMENT
AGENCY

Field Test Site

The TFID analyser was assessed on the basis of a four month trial mounted on a waste incinerator. H2 only fuel was used during the field test.

Approved Site Application

On the basis of these tests and the ranges for compliance with EU Directives this instrument is considered suitable for use on incineration and large combustion plant applications.

Any potential user should ensure, in consultation with the manufacturer, that the emission monitoring system is suitable for the process on which it will be installed. For general guidance on stack emission monitoring techniques refer to Environment Agency Technical Guidance Note M2: Monitoring of stack emissions to air. This is available on the Agency's website at www.environment-agency.gov.uk

Test Reports

This certification is based on the following Test Report(s) and on Sira's assessment and ongoing surveillance of the product and the manufacturing process:

TÜV Köln Report No: 936/809011/A dated 30.09.1999
TÜV Köln Report No: 936/807004/FID dated 26.02.1997

TÜV reports are accepted on the basis of the Environment Agency's document 'MCERTS – Guidance on the acceptance of German type approval test reports for CEMS' Version 2 (October 2003)

Product Certified

This certificate applies to instruments with software version 3.3x onwards.

Certificate No: Sira MC 050054/02
This Certificate Issued: 23 August 2011

This certificate may only be reproduced in its entirety and without change



ENVIRONMENT
AGENCY

Description

This instrument uses a Flame Ionisation Detector to measure hydrocarbon concentration. The sensor is in a heated chamber maintained at 120°C to 210°C. A regulated flow of sample gas passes through the flame sustained by fuel gas and air. Within the flame, hydrocarbons in the sample are ionised producing electrons and positive ions. Electrodes collect these ions, causing current to flow through an electronic measuring circuit proportional to the rate at which carbon atoms enter the burner

TFID may be configured as:

- Standalone Analyser: Comprises an analyser sensor, associated electronics and internal sample and utility gas transport components integrated into a housing, complete with internal power supply, control functionality, display, operator interface and input/output options – analogue, serial or digital.
- Analyser Module (AM): A blind analyser unit capable of measuring concentration, comprising detector sensor, associated electronics, internal sample and utility gas transport components. Concentration and other relevant data is combined and made available on a digital data highway to an Emerson NGA network or customer network.
- Integrated Network of Analysers: AM versions can be integrated in an NGA analyser system (mounted into a platform or combined with an MLT/TFID analyser or platform).

Mounting

The TFID is mounted in a standard 3U high, 19 inch rack mounting enclosure.

Sample Conditions

The TFID Analyser requires extractive sample conditioning equipment. The sample gas must be clean, dry, non-condensing and at a pressure of 800-1600mbar. The required sample flow rate is 0.5 to 1.5 l/min.

Certificate No: Sira MC 050054/02
This Certificate Issued: 23 August 2011



ENVIRONMENT
AGENCY

General Notes

1. This certificate is based upon the equipment tested. The Manufacturer is responsible for ensuring that on-going production complies with the standard(s) and performance criteria defined in this Certificate. The Manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management system shall be subject to regular surveillance according to 'Regulations Applicable to the Holders of Sira Certificates'. The design of the product certified is defined in the Sira Design Schedule for certificate No. Sira MC050054/01.
2. If certified product is found not to comply, Sira Certification Service should be notified immediately at the address shown on this certificate.
3. The Certification Marks that can be applied to the product or used in publicity material are defined in 'Regulations Applicable to the Holders of Sira Certificates'.
4. This document remains the property of Sira and shall be returned when requested by the company.

Certificate No: Sira MC 050054/02
This Certificate Issued: 23 August 2011