



PRODUCT CONFORMITY CERTIFICATE

This is to certify that the

9841 & 2041 NO_x Analyser

manufactured by:

Casella Monitor

*Regent House
Wolseley Road
Bedford
MK42 7JY*

has been assessed by Sira Certification Service
and for the conditions stated on this certificate complies with:

MCERTS Performance Standards for Continuous Ambient Air Quality Monitoring Systems, Version 7, dated October 2010

Certification Range :

NO	0 to 1200 µg/m ³ (0 - 962 nmol/mol)
NO ₂	0 to 500 µg/m ³ (0 - 261 nmol/mol)

Project No:	674/0165B
Certificate No:	Sira MC 070103/08
Initial Certification:	15 March 2007
This Certificate Issued	07 January 2011
Renewal Date:	14 March 2012

Technical Director

MCERTS is operated on behalf of the Environment Agency by

Sira Certification Service

12 Acorn Industrial Park, Crayford Road, Crayford
Dartford, Kent, UK, DA1 4AL
Tel: 01322 520500 Fax: 01322 520501

*This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please see www.siraenvironmental.com*



Approved Site Application

Any potential user should ensure, in consultation with the manufacturer, that the monitoring system is suitable for the intended application. For general guidance on monitoring techniques refer to the Environment Agency Monitoring Technical Guidance Notes available at www.mcerts.net

All tests have been conducted in accordance with BS EN 14211. On the basis of these tests this certificate is valid when the instrument is used for rural or urban air quality monitoring and similar applications; and in dilution systems where the sample concentration delivered to the analyser is within the certification range.

Basis of Certification

This certification is based on the following Test Report(s) and on Sira's assessment and ongoing surveillance of the product and the manufacturing process:

Umwelt Bundes Amt Frankfurt
US EPA Report
Sira

Report No. 428/96 dated 19/03/96
dated June 1992
Report Number 16A24048 dated 06/01/2010

Product Certified

This certificate applies to all instruments described by part numbers 98417000-1, 98417000-2, 98413000-104, 98413000-105 and 98415200-100 manufactured from 01 January 2006 onwards (serial number M2306-00 onwards and software version B1.27.2 and 3.19 onwards).

All 2041 instruments described by part numbers 204XXXC, 204XXXC (where 'XXX' are model options) manufactured from 01 December 2010 onwards. (serial number 4701925 onwards and software version 1.703 onwards)

Certificate No: Sira MC 070103/08
This Certificate Issued: 07 January 2011

*This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please see www.siraenvironmental.com*



Certified Performance

The instrument was evaluated for use under the following conditions:

Ambient Temperature Range: +5 to +40°C

Unless otherwise stated the evaluation was carried out on the certification range 0 to 2000 ppb.

Test	Results expressed as % of measured value				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Repeatability at zero					0.072 nmol/mol	<1 nmol/mol
Repeatability at hourly limit value					2.5 nmol/mol	<3 nmol/mol
Residual lack of fit at zero					4.5 nmol/mol	<5 nmol/mol
Lack of fit (largest residual from the linear regression line)				0.31		<4%
Sensitivity coefficient to sample gas pressure					0.15 nmol/mol/kPa	<8 nmol/mol/kPa
Sensitivity coefficient to sample gas temperature					No effect observed	<3 nmol/mol/K
Sensitivity coefficient to surrounding air temperature					Span: 0.36 nmol/mol/K	<3 nmol/mol/K
Sensitivity coefficient to electrical supply voltage					No effect observed	<0.3 nmol/mol/V
Converter efficiency					98%	>95%
Interference by H ₂ O (at concentration of 19 nmol/mol)					1.0 nmol/mol	<5 nmol/mol

Certificate No: Sira MC 070103/08

This Certificate Issued: 07 January 2011

*This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please see www.siraenvironmental.com*



Test	Results expressed as % of measured value				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Interference by NH ₃ (at concentration of 200 nmol/mol)					0.29 nmol/mol	<5 nmol/mol
Interference by CO ₂ at concentration of 500µmol/mol					0.02 nmol/mol	<5 nmol/mol
Interference by Ozone at concentration of 200 nmol/mol					0.20 nmol/mol	<2 nmol/mol
Averaging effect				3.08		<7%
Short term zero drift (over 12h)					0.06 nmol/mol	<2 nmol/mol
Short term span drift (over 12h)					0.30 nmol/mol	<6 nmol/mol
Response time (rise)					91 s	180 s
Response time (fall)					97 s	180 s
Difference between rise and fall time				3.6		<10%
Reproducibility under field conditions Note 1				2.65		<5% averaged over three month period
Long term zero drift (over 3 months) Note 1					0.05 nmol/mol	<5 nmol/mol
Long term span drift (over 3 months) Note 1				4.25		<5% of the max of certification range
Period of unattended operation Note 1					30 days	3 months not less than 2 weeks
Availability (data capture) Note 1					98%	>90%
Combined performance characteristic					12.37%	<15%

Note 1: The field test was performed on an urban site for 3 months.

Certificate No: Sira MC 070103/08
 This Certificate Issued: 07 January 2011

*This certificate may only be reproduced in its entirety and without change
 To authenticate the validity of this certificate please see www.siraenvironmental.com*



Description:

The 9841 & 2041 Chemiluminescence analyser consists of a low volume single reaction chamber and single detector. A photomultiplier tube measures the chemiluminescence produced when NO is mixed with Ozone. The air sample is fed via a flow block and bypass loop to ensure that the analyser measures Nitrogen Dioxide accurately in fast changing environments such as roadside applications. The flow block also allows a reference for the measurement system to be checked every 69 seconds providing additional stability. A microprocessor is used for controlling the various temperature zones and control loops and in addition compensates for temperature and pressure fluctuations. The microprocessor also stores a minimum of 1 years worth of 15 minute time and status stamped data for NO/NO_x and NO₂ and provides the facility of online remote diagnostics allowing all analyser functions to be controlled from a PC device such as a PDA or desk top PC. The 9841 & 2041 analyser employs a 'KALMAN' adaptive time averaging filter that gives the analyser fast response capability without creating inaccuracies due to fixed averaging.

General Notes

1. This certificate is based upon the equipment tested. The Manufacturer is responsible for ensuring that on-going production complies with the standard(s) and performance criteria defined in this Certificate. The Manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management system shall be subject to regular surveillance according to 'Regulations Applicable to the Holders of Sira Certificates'. The design of the product certified is defined in the Sira Design Schedule for certificate No. Sira MC 070103/05.
2. If certified product is found not to comply, Sira Certification Service should be notified immediately at the address shown on this certificate.
3. The Certification Marks that can be applied to the product or used in publicity material are defined in 'Regulations Applicable to the Holders of Sira Certificates'.
4. This document remains the property of Sira and shall be returned when requested by the company.

Certificate No: Sira MC 070103/08
This Certificate Issued: 07 January 2011

*This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please see www.siraenvironmental.com*