

# PRODUCT CONFORMITY CERTIFICATE

This is to certify that the

***TEOM 1405-F PM<sub>10</sub> FDMS***

manufactured by:

***Thermo Fisher Scientific***

*27 Forge Parkway  
Franklin,  
MA 02038  
USA*

has been assessed by Sira Certification Service  
and for the conditions stated on this certificate complies with:

**MCERTS Performance Standards for Continuous Ambient Air Quality Monitoring  
Systems, Version 8, dated June 2012**

**MCERTS for UK Particulate Matter as set out in the Annex to the MCERTS Performance  
Standards for Ambient Air Quality Monitoring Systems: Requirements of the UK  
Competent Authority for the Equivalence Testing and Certification of Automated  
Continuous and Manual Discontinuous Methods that Monitor Particulate Matter in  
Ambient Air, Version 1.1, dated 31 July 2012**

Certification Ranges:

PM<sub>10</sub>                      0 to 1000 µg/m<sup>3</sup>

Project No:                      16A28522 Add A  
Certificate No:                 Sira MC130210/01  
Initial Certification:         11 June 2013  
This Certificate Issued       18 May 2015  
Renewal Date:                 10 June 2018

R Cooper I Eng MInst MC  
Technical Director

MCERTS is operated on behalf of the Environment Agency by

## **Sira Certification Service**

12 Acorn Industrial Park, Crayford Road, Crayford  
Dartford, Kent, UK DA1 4AL  
Tel: +44 (0)1322 520500 Fax: +44 (0)1322 520501



*This certificate may only be reproduced in its entirety and without change*

**Registered Office: Rake Lane, Eccleston, Chester, UK CH4 9JN**

*To authenticate the validity of this certificate please visit [www.siracertification.com/mcerts](http://www.siracertification.com/mcerts)*

## Approved Site Application

Any potential user should ensure, in consultation with the manufacturer, that the monitoring system is suitable for the intended application. For general guidance on monitoring techniques refer to the Environment Agency Monitoring Technical Guidance Notes available at [www.mcerts.net](http://www.mcerts.net)

The PM<sub>10</sub> field test was conducted at one site in the UK and two sites in Germany. The particulate loading at the test sites is representative of different types of areas including urban background and rural areas affected by motorway traffic. The testing took place in both winter and summer months.

On the basis of these tests this certificate is valid when the instrument is used for urban and rural air quality monitoring and similar applications.

The requirements according to [Guide To The Demonstration Of Equivalence Of Ambient Air Monitoring Methods](#) (GDE 2010) are fulfilled for PM<sub>10</sub>.

## Basis of Certification

This certification is based on the following report(s) and on Sira's assessment and ongoing surveillance of the product and the manufacturing process:

MCERTS certification committee Report

[Certification Report and Checklist on the Evaluation of the Ambient Air Particulate Matter Test Reports Submitted for Approval and Certification within the MCERTS Scheme for UK Particulate Matter, 5 June 2013](#)

Bureau Veritas

[UK Report on the Equivalence of the PM<sub>10</sub> TEOM 1405-F. Report ref AGGX5508189/BV/DH/2833 dated 5 June 2013](#)

TÜV Rheinland

Report on the suitability test of the ambient air quality measuring system TEOM 1405-F Ambient Particulate Monitor with PM<sub>10</sub> pre-separator of the company Thermo Fisher Scientific for the component PM<sub>10</sub>.

Report number 936/21209885/B dated 25 November 2011.

Report published on [www.qal1.de/en/hersteller/thermo.htm](http://www.qal1.de/en/hersteller/thermo.htm)

TÜV Rheinland

Notice in relation to software upgrade from 1.55 to 1.56 dated 6 July 2012.

Notice published on [www.qal1.de/en/hersteller/thermo.htm](http://www.qal1.de/en/hersteller/thermo.htm)

## Product Certified

The **TEOM 1405-F with PM<sub>10</sub> pre-separator** measuring system consists of the following parts:

- USEPA style PM<sub>10</sub> sampling inlet operating at 16.7 l min<sup>-1</sup>;
- Flow splitter with flows split to 3 and 13.7 l min<sup>-1</sup>;
- Sampling tubes;
- FDMS (Filter Dynamics Measurement Systems) tower;
- TEOM (Tapered Element Oscillating Microbalance) mass sensor;
- Mass Flow Controllers;
- Vacuum pump.

This certificate applies to all 1405F PM<sub>10</sub> instruments fitted with software version 1.56 (serial number 20006 onwards).

Certificate No: Sira MC130210/01  
This Certificate Issued: 18 May 2015

**Certified Performance**

**Ambient temperature range:** 8°C to 25°C

The data in this table relate to the instrument **with no correction** for slope and/or intercept. See Note 1.

Test	Results	MCERTS specification
Constancy of the sample volumetric flow	-0.096%	To remain constant within $\pm 3\%$ of the rated value
Tightness of the sampling system	2.0% Note 2	Leakage not to exceed 1% of the sampled volume
Between sampler/instrument uncertainty for the standard method <b>PM<sub>10</sub></b>		
Full data set	0.48 $\mu\text{g}/\text{m}^3$	$\leq 2 \mu\text{g}/\text{m}^3$
<30 $\mu\text{g}/\text{m}^3$	0.46 $\mu\text{g}/\text{m}^3$	Not specified
$\geq 30 \mu\text{g}/\text{m}^3$	0.55 $\mu\text{g}/\text{m}^3$	Not specified
Between sampler/instrument uncertainty for the candidate method <b>PM<sub>10</sub></b>		
Full data set	1.09 $\mu\text{g}/\text{m}^3$	$\leq 2.5 \mu\text{g}/\text{m}^3$
<30 $\mu\text{g}/\text{m}^3$	1.03 $\mu\text{g}/\text{m}^3$	$\leq 2.5 \mu\text{g}/\text{m}^3$
$\geq 30 \mu\text{g}/\text{m}^3$	1.35 $\mu\text{g}/\text{m}^3$	$\leq 2.5 \mu\text{g}/\text{m}^3$

Test	Results	MCERTS specification
<b>Expanded uncertainty calculated at 50 µg/m<sup>3</sup> for Instrument SN 20006</b>		
Full data set	9.01	≤25%
≤30 µg/m <sup>3</sup>	12.26	Not specified
≥30 µg/m <sup>3</sup>	9.36	≤25%
Individual sites	Note 3	
Teddington Summer	23.71	≤25%
Cologne Winter	6.23	≤25%
Bornheim Summer	20.10	≤25%
Teddington Winter	15.22	≤25%
<b>Expanded uncertainty calculated at 50 µg/m<sup>3</sup> for Instrument SN 20107</b>		
Full data set	9.07	≤25%
≤30 µg/m <sup>3</sup>	13.06	Not specified
≥30 µg/m <sup>3</sup>	11.60	≤25%
Individual sites		
Teddington Summer	17.19	≤25%
Cologne Winter	9.35	≤25%
Bornheim Summer	21.03	≤25%
Teddington Winter	14.07	≤25%
Maintenance interval <b>PM<sub>10</sub></b>	Four weeks	≥Two weeks
Data Availability ( <b>SN20006</b> )	94.5%	≥90%
Data Availability ( <b>SN20107</b> )	98.6%	≥90%
Number of UK Tests	2	≥1
Number of Reference Methods	2	≥1

Note 1: For the PM<sub>10</sub> data, as the intercept was statistically significantly different from zero based upon 2 standard deviations, test results were also subjected to correction coefficients for intercept. Expanded uncertainties for the Candidate Method were calculated for both uncorrected datasets as well as data that have been adjusted for intercept. The CM fulfills the relevant Data Quality Objective of EU Directive 2008/50/EC when used without correction, though the highest individual expanded uncertainty specified in GDE 2010 (required to be below 25 %) is reduced if correction for intercept by subtracting 0.395 µg/m<sup>3</sup> is employed. Intercept correction is not required in order to make the instrument equivalent, but it is essential that thorough and frequent ongoing QAQC procedures are employed (as prescribed in prEN12341:2013 and prCEN/TS16450:2013) including to precisely quantify analyser baseline performance and ensure the continued correct operation of the FDMS drier.

PM <sub>10</sub> 1405-F	Calculated slope of all paired data	Calculated intercept of all paired data	Expanded uncertainty of all paired data	Range of individual expanded uncertainties
Uncorrected data	0.994	0.395	8.4%	6.2% to 23.7%
Data corrected for intercept by subtracting 0.395	0.994	0.000	8.5%	6.6% to 22.2%

Note 2: The leak test procedure for the 1405-F and the 1405-DF is an internal manufacturer's procedure, implemented in the instruments in order to avoid serious damage to the instrument. The check on tightness must be performed using this internal procedure. All of the leak tests conducted passed the manufacturer's leak test specifications.

Note 3: A study of pollution climate by Bureau Veritas relevant to sites in the UK and Germany has demonstrated that in all cases the particulate geometric mean criteria are met and at least one site meets the lower threshold and higher threshold criterion for wind speed, ambient temperature, ambient dew point and semi volatile nitrate content. The pollution climate criteria are satisfied for all the equivalence tests.

Note 4: For the purposes of quality control of these monitors in the field, as with all PM instruments that are not the reference method, the 1405 series FDMSs should be calibrated on a test site at intervals against the gravimetric reference methods EN 12341 or EN 14907 as applicable, and as given in the recommendations of the GDE 2010 and prEN16450.

### Description:

The TEOM 1405-F PM<sub>10</sub> Ambient Particulate Monitor with FDMS is a mass measurement monitor that incorporates the Tapered Element Oscillating Microbalance (TEOM) to measure particulate matter mass concentrations continuously. The TEOM 1405-F Monitor draws ambient air through a sample filter for collection at a constant flow rate. The monitor continuously weighs the filter calculating near real-time mass concentrations.

The 1405-F PM<sub>10</sub> is used to sample PM<sub>10</sub> by using a 16.7 l min<sup>-1</sup> USEPA style PM<sub>10</sub> inlet. The FDMS unit automatically determines mass concentration measurements that account for both non-volatile and volatile PM components.

The sensor unit contains a mass measurement hardware system that monitors particles that continuously accumulate on the system's exchangeable TEOM filters. By maintaining a flow rate of 3 l min<sup>-1</sup> through the sample channel, and measuring the total mass accumulated on the TEOM filter, the device can calculate the mass concentration in near real-time.

### General Notes

1. This certificate is based upon the equipment tested. The Manufacturer is responsible for ensuring that on-going production complies with the standard(s) and performance criteria defined in this Certificate. The Manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management system shall be subject to regular surveillance according to 'Regulations Applicable to the Holders of Sira Certificates'. The design of the product certified is defined in the Sira Design Schedule for certificate No. Sira MC130210/00
2. If certified product is found not to comply, Sira Certification Service should be notified immediately at the address shown on this certificate.
3. The Certification Marks that can be applied to the product or used in publicity material are defined in 'Regulations Applicable to the Holders of Sira Certificates'.
4. This document remains the property of Sira and shall be returned when requested by the company.