

PRODUCT CONFORMITY CERTIFICATE

This is to certify that the

AQM 65

Manufactured by:

Aeroqual Ltd

460 Rosebank Road,
Avondale
Auckland 1026
New Zealand

has been assessed by Sira Certification Service
And for the conditions stated on this certificate complies with:

**MCERTS Performance Standards for Indicative Ambient Particulate
Monitors, Version 4 dated August 2017**

Certification Ranges :

PM₁₀ 0-150µg/m³

Project No.: 70060477
Certificate No: Sira MC160289/01
Initial Certification: 16 March 2016
This Certificate issued: 16 September 2019
Renewal Date: 15 March 2021



Emily Alexander
Environmental Project Engineer

MCERTS is operated on behalf of the Environment Agency by

Sira Certification Service

Unit 6, Hawarden Industrial Park
Hawarden, Deeside, CH5 3US
Tel: +44 (0)1244 670 900



sira
CERTIFICATION

*The MCERTS certificate consists of this document in its entirety.
For conditions of use, please consider all the information within.
This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Certificate Contents

Approved Site Application.....	2
Basis of Certification	2
Product Certified.....	2
Certified Performance	3
Description.....	4
General Notes	4

Approved Site Application

Any potential user should ensure, in consultation with the manufacturer, that the monitoring system is suitable for the intended application. For general guidance on monitoring techniques refer to the Environment Agency Monitoring Technical Guidance Notes available at www.mcerts.net

The indicative dust monitoring analyser(s) can be operated in one of two ways:

For qualitative measurements: Providing qualitative measurement data for the analysis of particulate pollution trends, and source identification studies based for example on pollution roses etc. Such application can rely on instrument factory calibration only.

For quantitative measurements: Providing measurement data with the uncertainty defined for indicative instruments (+/- 50%). This can be achieved on condition that each instrument used for measurement has been calibrated on the specific site where monitoring is taking place against a standard reference method for a period of two weeks and the resulting slope and intercept have been used for instrument calibration. Using non-standard filters and procedures for this purpose is not acceptable. To maintain the validity of data this calibration has to be repeated at least every twelve months or when the instrument is moved to a different site.

They **cannot** be used as a substitute for continuous ambient air quality monitoring systems (CAMs) employed in national air quality monitoring networks for the EU Air Quality Directive

Basis of Certification

This certification is based on the following Test Report(s) and on Sira's assessment and ongoing surveillance of the product and the manufacturing process:

BECA Report Aeroqual M Certification Requirements dated 31/07/2013
Sira Report 16A27887 dated 25/09/2013

Product Certified

The measuring system consists of the following parts:

- AQM 65
- AQM PM₁₀ Cyclone

This certificate applies to all instruments fitted with software version 1.2 and serial number -406 onwards.

Certificate No : Sira MC160289/01
This Certificate issued : 16 September 2019

*This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Certified Performance

Test	Result	MCERTS specification
Constancy of the sample volumetric flow	<±3%	Remain constant within ±3% of rated value
Tightness of the sampling system	0.9%	Leakage not to exceed 2% of sampled volume
Intra-instrument uncertainty for the reference method All data ≥ 30 µg/m ³ ≤ 30 µg/m ³	0.46 µg/m ³ 0.43 µg/m ³ 0.46 µg/m ³	≤5µg/m ³
Intra-instrument uncertainty for the candidate method All data ≥ 30 µg/m ³ ≤ 30 µg/m ³	1.39 µg/m ³ 2.21 µg/m ³ 1.18 µg/m ³	≤5µg/m ³
Highest resulting uncertainty estimate comparison against data quality objective (Measurement Uncertainty)	23.5%	WCM≤Wd _{qo} Measurement uncertainty defined as 50% for indicative instruments
Maintenance Interval	Two weeks	>Two weeks

Certificate No : Sira MC160289/01
 This Certificate issued : 16 September 2019

*This certificate may only be reproduced in its entirety and without change
 To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Description

The AQM65 employs a near forward light scattering nephelometer and PM₁₀ sharp cut cyclone to measure PM₁₀.

The nephelometer uses a collimated beam of light from a laser operating at 670 nm to illuminate the incoming sample air. Light scattered by particles in the air is collected and focused on to a photodiode which converts the light intensity to an electrical signal. The signal is amplified and calibrated to provide an output in terms of particulate mass. The nephelometer has an on-board temperature sensor which corrects for thermal drift and sheath air to keep the optics clean.

Automatic baseline drift correction is achieved using a regular air purge cycle which pushes filtered air into the nephelometer to enable a particle-free background to be measured and subtracted from the mass signal. A 10 W inlet heater on the sample inlet tube reduces the humidity of sampled air to prevent particle growth and fogging of the nephelometer optics.

A diaphragm pump is used to provide a 2.0 LPM sample flow which is matched to the sharp cut cyclone design to ensure particles with diameters larger than 10 micron are removed.

General Notes

1. This certificate is based upon the equipment tested. The Manufacturer is responsible for ensuring that on-going production complies with the standard(s) and performance criteria defined in this Certificate. The Manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management system shall be subject to regular surveillance according to 'Regulations Applicable to the Holders of Sira Certificates'.
2. The design of the product certified is defined in the Sira Design Schedule V1.1 for certificate No. Sira MC160289/01
3. If certified product is found not to comply, Sira Certification Service should be notified immediately at the address shown on this certificate.
3. The Certification Marks that can be applied to the product or used in publicity material are defined in 'Regulations Applicable to the Holders of Sira Certificates'.
4. This document remains the property of Sira and shall be returned when requested by the company.

Certificate No : Sira MC160289/01
This Certificate issued : 16 September 2019

*This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*