

PRODUCT CONFORMITY CERTIFICATE

This is to certify that the

AR 602Z (UV) & AR 650 (IR)
used either independently or in combination

manufactured by:

Opsis AB
P.O. Box 244
S-244 02 Furulund
Sweden


has been assessed by Sira Certification Service
and for the conditions stated on this certificate complies with:

Environment Agency Guidance
“MCERTS for stack emissions monitoring equipment at industrial installations”
- Continuous emissions monitoring systems (CEMS)
Published 20 October 2020
EN 15267-1 :2009, EN 15267-2 :2009, EN 15267-3 :2007
& QAL 1 as defined in EN 14181: 2014

Certification ranges:

	AR 650 (IR)		AR 602Z (UV)
CO	- 0 to 75 mg/m ³	SO ₂	- 0 to 75 mg/m ³
HCl	- 0 to 15 mg/m ³	NO ₂	- 0 to 20 mg/m ³
H ₂ O	- 0 to 30 %vol.	NH ₃	- 0 to 10 mg/m ³
HF	- 0 to 5 mg/m ³	NO	- 0 to 150 mg/m ³
		Formaldehyde	- 0 to 20 mg/m ³
		Phenol	- 0 to 20 mg/m ³
		H ₂ O	- 0 to 30 mg/m ³

Project number: 80070051
Certificate number: Sira MC020011/08
Initial certification: 01 March 2002
This certificate issued: 15 March 2021
Renewal date: 28 March 2026



Andrew Young
Environmental Team Manager

MCERTS is operated on behalf of the Environment Agency by

Sira Certification Service

Unit 6, Hawarden Industrial Park
Hawarden, Deeside, CH5 3US
Tel: +44 (0)1244 670 900



0011



*The MCERTS certificate consists of this document in its entirety.
For conditions of use, please consider all the information within.
This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Certificate Contents

Approved Site Application.....	2
Basis of Certification	2
Product Certified.....	3
Certified Performance	4
Description.....	15
General Notes	15

Approved Site Application

Any potential user should ensure, in consultation with the manufacturer, that the monitoring system is suitable for the intended application. For general guidance on monitoring techniques refer to the Environment Agency Monitoring Technical Guidance Notes available at www.mcerts.net

This instrument is considered suitable for use on waste incineration and large combustion plant applications. This CEMS has been proven suitable for its measuring task (parameter and composition of the flue gas) by use of the QAL 1 procedure specified in EN14181. The lowest certified range for each determinand shall not be more than 1.5 times the daily average emission limit value (ELV) for incineration plants, and not more than 2.5 times the ELV for other types of application.

This certification is based on the following Test Report(s) and on Sira's assessment and ongoing surveillance of the product and the manufacturing process:

Basis of Certification

Original certification:

TÜV Rheinland report 936/804002/NH₃ dated 06/06/94
TÜV Rheinland report 936/800010 dated 27/06/92
TÜV Rheinland report 936/804001 from April 26, 1996 (CO, HCl and H₂O)
TÜV Rheinland report 936/21201391/A from June 30, 2004 (HF)
TÜV Rheinland report 936/80009 from August 02, 1991 (SO₂, NO and NO₂)
TÜV Rheinland report 936/806013 from August 02, 1991 (SO₂, NO and NO₂)
TÜV Rheinland report 936/807024/A from September 30, 1999 (SO₂, NO and NO₂)
TÜV Rheinland report 936/800010 from March 30, 1992 (NH₃)
TÜV Rheinland report 936/802011 from June 06, 1994 (Formaldehyde and Phenol)
TÜV Rheinland report 936/807017 from March 12, 1998 (Formaldehyde)
TÜV Rheinland report 936/800010/2 from March 01, 1993 (H₂O)
Sira report N0393 dated Feb 2002

Recertification:

TÜV Rheinland report 936/21213004/A dated 30/11/2010
TÜV Rheinland report 936/21213004/C dated 30/11/2010
Sira Evaluation Report 16A24051 (AR 650) dated 21/01/2011
Sira Evaluation Report 16A24051 (AR 602Z) dated 21/01/2011

Certificate number: Sira MC020011/08
This certificate issued: 15 March 2021

Product Certified

The AR650/602Z measuring system consists of the following parts:

- Receiver unit (Model RE062)
- Transmitter unit (Model EM 062-A)
- Control unit (analyser)

This certificate applies to all instruments fitted with software version 7.21 onwards (serial number 160 onwards).

Certificate number: Sira MC020011/08
This certificate issued: 15 March 2021

*This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Certified Performance

The instrument was evaluated for use under the following conditions:

Ambient Temperature Range: Stack components -30°C to +60°C
Control unit +5°C to +35°C

IP rating: IP20: Analyser, must be placed in a protected area
IP54: Duct mounted parts (transmitter & receiver unit)

Unless otherwise stated the evaluation was carried out on the certification ranges, CO 0 to 75mg/m³, HCl 0 to 15mg/m³, H₂O 0 to 30 %vol. (IR), HF 0 to 5 mg/m³, SO₂ 0 to 75mg/m³, NO₂ 0 to 20mg/m³, NH₃ 0 to 10mg/m³, NO 0 to 150mg/m³, Formaldehyde 0 to 20 mg/m³, Phenol 0 to 20 mg/m³, H₂O 0 to 30 mg/m³ (UV).

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Response time						
CO					<120s	<200s
HCl					<120s	<400s
H ₂ O (30 %vol.)					<120s	<200s
HF					<120s	<400s
SO ₂					<180s	<200s
NO ₂					<180s	<200s
NH ₃					<180s	<400s
NO					<180s	<200s
Formaldehyde					<180s	<200s
Phenol					<180s	<200s
H ₂ O (30 mg/m ³)					<180s	<200s
Repeatability standard deviation at zero point						
CO		0.8				<2.0%
HCl			1.8			<2.0%
H ₂ O (30 %vol.)	0.1					<2.0%
HF	0.8					<2.0%
SO ₂	0.1					<2.0%
NO ₂	0.3					<2.0%

Certificate number: Sira MC020011/08
This certificate issued: 15 March 2021

*This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
NH ₃		0.8				<2.0%
NO	0.1					<2.0%
Formaldehyde	0.5					<2.0%
Phenol	0.1					<2.0%
H ₂ O (30 mg/m ³)	0.2					<2.0%
Repeatability standard deviation at reference point						
CO	0.4					<2.0%
HCl			1.3			<2.0%
H ₂ O (30 %vol.)	0.2					<2.0%
HF		0.9				<2.0%
SO ₂	0.1					<2.0%
NO ₂	0.2					<2.0%
NH ₃			1.2			<2.0%
NO	0.2					<2.0%
Formaldehyde	0.4					<2.0%
Phenol	0.5					<2.0%
H ₂ O (30 mg/m ³)	0.3					<2.0%
Lack-of-fit						
CO		0.93				<2.0%
HCl			2.00			<2.0%
H ₂ O (30 %vol.)		0.67				<2.0%
HF			-1.80			<2.0%
SO ₂		0.63				<2.0%
NO ₂		0.70				<2.0%
NH ₃		0.70				<2.0%
NO		-0.73				<2.0%
Formaldehyde		1.00				<2.0%

Certificate number: Sira MC020011/08
 This certificate issued: 15 March 2021

*This certificate may only be reproduced in its entirety and without change
 To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Phenol	0.50					<2.0%
H ₂ O (30 mg/m ³)		1.00				<2.0%
Influence of ambient temperature zero point						
CO	-0.40					<5.0%
HCl	0.20					<5.0%
H ₂ O (30 %vol.)	0.00					<5.0%
HF			-1.20			<5.0%
SO ₂	-0.20					<5.0%
NO ₂	0.20					<5.0%
NH ₃	0.10					<5.0%
NO	-0.10					<5.0%
Formaldehyde	-0.40					<5.0%
Phenol	0.10					<5.0%
H ₂ O (30 mg/m ³)	0.20					<5.0%
Influence of ambient temperature reference point						
CO			-1.20			<5.0%
HCl	0.20					<5.0%
H ₂ O (30 %vol.)	0.03					<5.0%
HF				4.00		<5.0%
SO ₂	0.50					<5.0%
NO ₂	-0.50					<5.0%
NH ₃		-1.00				<5.0%
NO	0.10					<5.0%
Formaldehyde	-0.50					<5.0%
Phenol	0.50					<5.0%
H ₂ O (30 mg/m ³)	0.30					<5.0%

Certificate number: Sira MC020011/08
 This certificate issued: 15 March 2021

*This certificate may only be reproduced in its entirety and without change
 To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Influence of sample gas pressure						
CO		-0.53				<2.0%
HCl		-0.89				<2.0%
H ₂ O (30 % ^{vol.})	-0.33					<2.0%
HF	-0.47					<2.0%
SO ₂	-0.49					<2.0%
NO ₂		-0.67				<2.0%
NH ₃			1.21			<2.0%
NO	0.33					<2.0%
Formaldehyde	-0.50					<2.0%
Phenol	-0.50					<2.0%
H ₂ O (30 mg/m ³)	0.11					<2.0%
Influence of voltage variations 190 to 250V						
CO	0.50					<2.0%
HCl		-0.90				<2.0%
H ₂ O (30 % ^{vol.})	-0.10					<2.0%
HF		-0.80				<2.0%
SO ₂	0.20					<2.0%
NO ₂	0.50					<2.0%
NH ₃		-1.00				<2.0%
NO	0.10					<2.0%
Formaldehyde		0.90				<2.0%
Phenol	0.20					<2.0%
H ₂ O (30 mg/m ³)		0.70				<2.0%
Influence of vibration (10 to 60Hz (±0.3mm), 60 to 150Hz at 19.6m/s ²)					Note 1	To be reported

Certificate number: Sira MC020011/08
 This certificate issued: 15 March 2021

*This certificate may only be reproduced in its entirety and without change
 To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Cross-sensitivity at zero with interferents: O ₂ , H ₂ O, CO, CO ₂ , CH ₄ , N ₂ O, NO, NO ₂ , NH ₃ , SO ₂ , HCl						
CO	-0.44					<4.0%
HCl	0.00					<4.0%
H ₂ O (30 % ^{vol.})	-0.67					<4.0%
HF	0.00					<4.0%
SO ₂			-1.21			<4.0%
NO ₂				2.60		<4.0%
NH ₃			2.00			<4.0%
NO					Note 2	<4.0%
Formaldehyde	0.40					<4.0%
Phenol			2.00			<4.0%
H ₂ O (30 mg/m ³)					Note 2	<4.0%
Cross-sensitivity at reference with interferents: O ₂ , H ₂ O, CO, CO ₂ , CH ₄ , N ₂ O, NO, NO ₂ , NH ₃ , SO ₂ , HCl						
CO	0.84					<4.0%
HCl	0.94					<4.0%
H ₂ O (30 % ^{vol.})	0.57					<4.0%
HF	0.00					<4.0%
SO ₂			-1.96			<4.0%
NO ₂				-2.85		<4.0%
NH ₃				2.30		<4.0%
NO					Note 2	<4.0%
Formaldehyde	0.50					<4.0%
Phenol			1.35			<4.0%
H ₂ O (30 mg/m ³)					Note 2	<4.0%

Certificate number: Sira MC020011/08
 This certificate issued: 15 March 2021

*This certificate may only be reproduced in its entirety and without change
 To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Excursion of measurement beam of cross-stack in-situ CEMS						
CO		0.93				<2.0%
HCl				1.33		<2.0%
H ₂ O (30 %vol.)		-0.85				<2.0%
HF			-1.25			<2.0%
SO ₂		-0.64				<2.0%
NO ₂			1.25			<2.0%
NH ₃			2.00			<2.0%
NO		-0.62				<2.0%
Formaldehyde			1.80			<2.0%
Phenol		0.70				<2.0%
H ₂ O (30 mg/m ³)		-0.77				<2.0%
Converter Efficiency					N/A	>95%

Certificate number: Sira MC020011/08
 This certificate issued: 15 March 2021

*This certificate may only be reproduced in its entirety and without change
 To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Measurement uncertainty						
CO					6.9%	7.5%
HCl					12.5%	30%
H ₂ O (30 % ^{vol.})					14.7%	22.5%
HF					2.6%	30%
SO ₂					6.7%	15%
NO ₂					6.3%	15%
NH ₃					24.5%	30%
NO					5%	15%
Formaldehyde					12.7%	22.5%
Phenol					6.4%	22.5%
H ₂ O (30 mg/m ³)					5.1%	7.5%
Calibration function (field)						
CO					0.99-0.97	>0.90
HCl					0.97-0.93	>0.90
H ₂ O (30 % ^{vol.})					1.00	>0.90
HF					1.00-0.90	>0.90
SO ₂					1.00	>0.90
NO ₂					1.00-0.99	>0.90
NH ₃					1.00-0.99	>0.90
NO					1.00-0.91	>0.90
Formaldehyde					0.97-0.95	>0.90
Phenol					0.99	>0.90
H ₂ O (30 mg/m ³)					0.97-0.90	>0.90

Certificate number: Sira MC020011/08
 This certificate issued: 15 March 2021

*This certificate may only be reproduced in its entirety and without change
 To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Response time (field)						
CO					<120s	<200s
HCl					<120s	<400s
H ₂ O (30 % ^{vol.})					<120s	<200s
HF					<120s	<400s
SO ₂					<180s	<200s
NO ₂					<180s	<200s
NH ₃					<180s	<400s
NO					<180s	<200s
Formaldehyde					<180s	<200s
Phenol					<180s	<200s
H ₂ O (30 mg/m ³)					<180s	<200s
Lack of fit (field)						
CO		-0.93				<2.0%
HCl		0.67				<2.0%
H ₂ O (30 % ^{vol.})			<2.00		Note 3	<2.0%
HF			<2.00		Note 3	<2.0%
SO ₂	-0.40					<2.0%
NO ₂			<2.00		Note 3	<2.0%
NH ₃			-1.30			<2.0%
NO		-0.60				<2.0%
Formaldehyde		1.00				<2.0%
Phenol			1.50			<2.0%
H ₂ O (30 mg/m ³)			<2.00		Note 3	<2.0%

Certificate number: Sira MC020011/08
 This certificate issued: 15 March 2021

*This certificate may only be reproduced in its entirety and without change
 To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Maintenance interval						>8 days
CO					6-months	>8 days
HCl					6-months	>8 days
H ₂ O (30 %vol.)					6-months	>8 days
HF					1-month	>8 days
SO ₂					6-months	>8 days
NO ₂					6-months	>8 days
NH ₃					6-months	>8 days
NO					3-months	>8 days
Formaldehyde					1-month	>8 days
Phenol					1-month	>8 days
H ₂ O (30 mg/m ³)					1-month	>8 days
Zero and Span drift requirement	The CEM does not contain an automatic correction of zero and span drift.					Clause 6.13 & 10.13 Manufacturer shall provide a description of the technique to determine and compensate for zero and span drift.

Certificate number: Sira MC020011/08
 This certificate issued: 15 March 2021

*This certificate may only be reproduced in its entirety and without change
 To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Change in zero point over maintenance interval						
CO	0.40					<3.0%
HCl		0.60				<3.0%
H ₂ O (30 %vol.)			2.00		Note 4	<3.0%
HF			2.00		Note 4	<3.0%
SO ₂	0.30				Note 4	<3.0%
NO ₂			1.50		Note 4	<3.0%
NH ₃			2.00		Note 4	<3.0%
NO	0.50				Note 4	<3.0%
Formaldehyde		1.00			Note 4	<3.0%
Phenol		1.00			Note 4	<3.0%
H ₂ O (30 mg/m ³)			1.50		Note 4	<3.0%
Change in reference point over maintenance interval						
CO		1.00				<3.0%
HCl		1.00				<3.0%
H ₂ O (30 %vol.)				3.00	Note 4	<3.0%
HF				3.00	Note 4	<3.0%
SO ₂			2.00		Note 4	<3.0%
NO ₂				3.00	Note 4	<3.0%
NH ₃				3.00	Note 4	<3.0%
NO			2.00		Note 4	<3.0%
Formaldehyde				2.30	Note 4	<3.0%
Phenol			2.00		Note 4	<3.0%
H ₂ O (30 mg/m ³)				3.00	Note 4	<3.0%
Availability					>95.4%	>95% (>98% for O ₂)

Certificate number: Sira MC020011/08
 This certificate issued: 15 March 2021

*This certificate may only be reproduced in its entirety and without change
 To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Reproducibility					
CO				2.10	<3.3%
HCl			1.20		<3.3%
H ₂ O (30 % ^{vol.})		0.70			<3.3%
HF		1.00			<3.3%
SO ₂				2.20	<3.3%
NO ₂			1.50		<3.3%
NH ₃			1.80		<3.3%
NO				2.90	<3.3%
Formaldehyde			1.70		<3.3%
Phenol				2.80	<3.3%
H ₂ O (30 mg/m ³)				2.30	<3.3%

Note 1: The instrument exhibited some moderate resonances. Some resonances caused the light source to go out. The effect was only temporary and the system functioned correctly once restored. The instrument did not appear to suffer any mechanical degradation (pre and post calibration).

Note 2: All deviations below 0.5% are considered to be negligible and not reported.

Note 3: Based on field calibration function test and laboratory lack of fit test. The lack of fit in the field must be verified during every check of installation of the CEM.

Note 4: Results taken from original report from initial certification, no additional testing required.

Certificate number: Sira MC020011/08
 This certificate issued: 15 March 2021

*This certificate may only be reproduced in its entirety and without change
 To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Description

The OPSIS CEM system is a cross stack flue gas measurement system that uses either an AR602Z (UV) analyser or a AR650 (IR) or a combination of both depending on the range of pollutants to be measured. A combination system was tested for MCERTS.

The AR 602Z is based upon UV absorption techniques for measuring SO₂, NO and NO₂, H₂O, HCl, NH₃, Hg and CO₂.

The AR 650 is based upon IR absorption techniques for measuring CO, HCl, HF and H₂O. The transmitter and receiver units are mounted opposite each other on a duct or stack. The receiver is connected to the control units by fibre optic cable. A common transmitter/receiver assembly is used with both control units.

This certificate applies to the AR602Z (UV) and AR650 (IR) either individually or in combination for the gases listed on page 1 of this certificate only.

Other gases can be measured with this system but were not tested as part of MCERTS.

General Notes

1. This certificate is based upon the equipment tested. The Manufacturer is responsible for ensuring that on-going production complies with the standard(s) and performance criteria defined in this Certificate. The Manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management system shall be subject to regular surveillance according to 'Regulations Applicable to the Holders of Sira Certificates'.
2. The design of the product certified is defined in the Sira Design Schedule for certificate No. Sira MC 020011/07
3. If certified product is found not to comply, Sira Certification Service should be notified immediately at the address shown on this certificate.
3. The Certification Marks that can be applied to the product or used in publicity material are defined in 'Regulations Applicable to the Holders of Sira Certificates'.
4. This document remains the property of Sira and shall be returned when requested by the company.

Certificate number: Sira MC020011/08
This certificate issued: 15 March 2021