

PRODUCT CONFORMITY CERTIFICATE

This is to certify that the

***IQ SENSOR NET 2020XT, 2020XT-H3 and 2020XT-H3-C6
Controller Unit
FDO 700 IQ Optical Dissolved Oxygen Probe
VisoTurb 700 IQ Turbidity Probe***

Manufactured by:

Xylem Analytics Germany GmbH
Dr.-Karl-Slevogt-Straße 1
D-82362 Weilheim
Germany

has been assessed by Sira Certification Service
And for the conditions stated on this certificate complies with:

**MCERTS Performance Standards for Continuous Water Monitoring Equipment,
Part 2: online analysers, Version 3.1 dated August 2010**

Certification Ranges :

DO	0 to 20mg/L / 0 to 200% sat
Turbidity	0 to 100 NTU 0 to 500 NTU

Project No.: 16W31523
Certificate No: Sira MC150281/01
Initial Certification: 21 October 2015
This Certificate issued: 01 July 2016
Renewal Date: 21 October 2020

Joe Prince MSc, MInst MC
Deputy Certification Manager

MCERTS is operated on behalf of the Environment Agency by

Sira Certification Service

Unit 6, Hawarden Industrial Park
Hawarden, Deeside, CH5 3US
Tel: +44 (0)1244 670 900



*The MCERTS certificate consists of this document in its entirety.
For conditions of use, please consider all the information within.
This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Certificate Contents

Approved Site Application	2
Basis of Certification	2
Product Certified.....	2
Certified Performance	3
Description.....	5
General Notes	8

Approved Site Application

Any potential user should ensure, in consultation with the manufacturer, that the monitoring system is suitable for the intended application. For general guidance on monitoring techniques refer to the Environment Agency Monitoring Technical Guidance Notes available at www.mcerts.net

On the basis of the assessment this instrument is considered suitable for use on treated wastewater, untreated wastewater and receiving water applications.

Field trial was conducted on an aeration lane of a wastewater treatment works for 3 months.

Basis of Certification

This certification is based on the following Test Report(s) and on Sira's assessment and ongoing surveillance of the product and the manufacturing process:

WRc report UC10125.01 October 2014
WRc report P9719 March 2014
WRc report UC10125.02 February 2015

Product Certified

The measuring system consists of the following parts:

- IQ SensorNet 2020XT, 202XT-H3 and 2020XT-H3-C6 Terminal/Controller
- FDO 700 IQ optical dissolved oxygen probe and/or VisoTurb 700 IQ turbidity sensor

This certificate applies to all instruments fitted with software version and serial number onwards.

- | | | | | |
|------------------|------------|----------|---------------|------|
| • MIQ/TC 2020 XT | Serial No: | 08330014 | Software Ver: | 3.59 |
| • FDO 700 IQ | Serial No: | 06440003 | Software Ver: | 2.27 |
| • VisoTurb IQ | Serial No: | 04210008 | Software Ver: | 2.54 |

Certificate No : Sira MC150281/01
This Certificate issued : 01 July 2016

*This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Certified Performance

The instrument was evaluated for use under the following conditions:

Ambient Temperature Range: -20°C to 55°C

Unless otherwise stated the evaluation was carried out on the certification ranges: DO 0 to 20mg/L, turbidity 0 to 100 NTU

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Combined performance characteristic						
DO				3.28		6% reading
Turbidity				2.31		2.5% span
Warm up time						
DO					75s	Value to be reported
Turbidity					71s	
Response time						
DO					270 seconds	Value to be reported
Turbidity					3 seconds	
Mean error						
DO			-1.91			5% reading
Turbidity (0 to 100 NTU)			1.8			2% span
Turbidity (0 to 500 NTU)			1.3			2% span
Linearity						
DO			1.23			2.5% reading
Turbidity (0 to 100 NTU)	-0.18					1% span
Turbidity (0 to 500 NTU)	0.35					

Certificate No : Sira MC150281/01
This Certificate issued : 01 July 2016

*This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Repeatability						
DO				2.18		2.5% reading
Turbidity (0 to 100 NTU)	0.09					1% span
Turbidity (0 to 500 NTU)	0.25					1% span
Sample matrix effects						
DO (salinity compensation)		0.6				2.5% reading
Turbidity (colour effects)			1.75			-
Drift						
DO			1.1			2.5% reading
Turbidity		-0.54				1% span
Output impedance (10 to 500Ω)						
DO					No effect	1% reading
Turbidity					No effect	0.5% span
Supply voltage (100V to 240V)						
DO	0.05					1% reading
Turbidity	0.125					0.5% span
Supply voltage DC (21.01V to 28.01V)						
DO	-0.4					1% reading
Turbidity	0.5					0.5% span
Ambient temperature (-20°C to +55°C)						
DO		0.68				2.5% reading
Turbidity		0.59				1% span

Certificate No : Sira MC150281/01
This Certificate issued : 01 July 2016

*This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Relative humidity (95% RH)						
DO	0.30					2.5% reading
Turbidity	0.03					1% span
Incident light						
DO	-0.24					1% reading
Turbidity	0.00					1% span
Sample temperature (3°C to 30°C)						
DO		0.73				2.5% reading
Turbidity	0.11					1% span
Sample flow rate						
DO	0.14					1% reading
Turbidity	0.25					0.5% span

Certificate No : Sira MC150281/01
 This Certificate issued : 01 July 2016

*This certificate may only be reproduced in its entirety and without change
 To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Field test results

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Error under field conditions						
DO					90%	>90% of errors
Turbidity					100%	≤ Uc value
Response time (start)						
DO					10 minutes	To be reported
Turbidity					58 seconds	
Up time					100%	>95%
Maintenance	None					To be reported

Certificate No : Sira MC150281/01
 This Certificate issued : 01 July 2016





*This certificate may only be reproduced in its entirety and without change
 To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Description

IQ SENSOR NET 2020XT, 202XT-H3 and 2020XT-H3-C6 Terminal/Controller Unit

The 2020XT is a digital terminal / controller which allows up to 20 IQ sensors to be installed on a single networked system to provide a complete monitoring solution from inlet to final effluent.

The IQ SensorNet system is broken down into 4 simple components that can be configured into configurations to tailor monitoring requirements and locations in a wastewater treatment plant. These components consist of the following:

Terminal / Controller	Modules	Cables	Sensors
			

All components can be connected together with power and communication for the entire network can provided on a single Xylem cable.

- Removable screen
- Up to 20 digital sensors
- Modular design
- Centralised power supply along entire network
- Analogue and digital outputs including PROFIBUS, MODBUS, RS-232
- LED status update
- GPRS or LAN connectivity

FDO 700 IQ

The FDO is an optical, luminescent based DO sensor for the IQ SensorNet system. During the biological nutrient removal process at wastewater treatment plants, continuous and precise measurement of dissolved oxygen concentration is of vital importance for optimal performance and trouble-free operation.

VisoTurb 700 IQ

The VisoTurb is a part of a family of optical sensors for turbidity and suspended solids measurements. These sensors incorporate an ultrasonic cleaning system. Turbidity measurements are carried out using a nephelometric principle. The best resolution for each measured value can be selected using an autorange function. Thus, one sensor is required to cover applications ranging from drinking water to highly concentrated sludge.

Certificate No : Sirra MC150281/01
This Certificate issued : 01 July 2016

*This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

General Notes

1. This certificate is based upon the equipment tested. The Manufacturer is responsible for ensuring that on-going production complies with the standard(s) and performance criteria defined in this Certificate. The Manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management system shall be subject to regular surveillance according to 'Regulations Applicable to the Holders of Sira Certificates'. The design of the product certified is defined in the Sira Design Schedule V00 for certificate No. Sira MC150281/01
2. If certified product is found not to comply, Sira Certification Service should be notified immediately at the address shown on this certificate.
3. The Certification Marks that can be applied to the product or used in publicity material are defined in 'Regulations Applicable to the Holders of Sira Certificates'.
4. This document remains the property of Sira and shall be returned when requested by the company.

Certificate No : Sira MC150281/01
This Certificate issued : 01 July 2016

*This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*