

PRODUCT CONFORMITY CERTIFICATE

This is to certify that the

Servomex 4900 Multigas Analyser

Manufactured by:

Servomex Group Ltd

Jarvis Brook
Crowborough
East Sussex
TN6 3DU
UK

Technical and Service Center
4 Constitution Way
Woburn
MA 01801-1087
USA

has been assessed by Sira Certification Service
and for the conditions stated on this certificate complies with:

**MCERTS Performance Standards for Continuous Emission
Monitoring Systems, Version 3.5 dated June 2016
EN15267-3:2007 & QAL 1
as defined in EN 14181: 2014**

Certification Ranges :

SO ₂	0 to 572 mg/m ³	CO	0 to 75 mg/m ³
NO	0 to 268 mg/m ³	O ₂	0 to 25 % vol.

Certification is awarded in respect of the conditions stated in this certificate

Project No. : 674/0336
Certificate No: Sira MC030013/11
Initial Certification: 03 July 2003
This Certificate issued : 06 March 2018
Renewal Date: 05 March 2023

Emily Alexander BSc (Hons)
Deputy Certification Manager

MCERTS is operated on behalf of the Environment Agency by

Sira Certification Service

Unit 6, Hawarden Industrial Park
Hawarden, Deeside, CH5 3US
Tel: +44 (0)1244 670 900



*The MCERTS certificate consists of this document in its entirety.
For conditions of use, please consider all the information within.
This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Certificate Contents

Approved Site Application.....	2
Basis of Certification	2
Product Certified.....	3
Certified Performance	4
Description.....	9
General Notes	9

Approved Site Application

Any potential user should ensure, in consultation with the manufacturer, that the monitoring system is suitable for the intended application. For general guidance on monitoring techniques refer to the Environment Agency Monitoring Technical Guidance Notes available at www.mcerts.net

On the basis of the assessment and the ranges required for compliance with EU Directives this instrument is considered suitable for use on waste incineration and large coal-fired combustion plant applications. This CEM has been proven suitable for its measuring task (parameter and composition of the flue gas) by use of the QAL 1 procedure specified in EN14181, for IED Chapter III and IED Chapter IV applications for the ranges specified. The lowest certified range for each determinand shall not be more than 1.5X the daily average emission limit value (ELV) for IED Chapter IV applications, and not more than 2.5X the ELV for IED Chapter III and other types of application.

The manufacturer states that the Servomex 4900 is not suitable for use with corrosive samples and consequently must always be used with an appropriate sample system. Potential interference's are site specific and may vary from stack to stack.

Basis of Certification

This certification is based on the following Test Report(s) and on Sira's assessment and ongoing surveillance of the product and the manufacturing process:


TÜV Essen Report:	Report number RWTÜV-3.5.2/0784/95//674377/01 dated 1997
	Report number RWTÜV-3.5.2/0784/95//597632/01 dated 1996
	Report number RWTÜV-5.0.2/0784/94//20363886 dated 1999
SIRA Report:	Report number N 0415 dated November 2002
TÜV Köln Report	Report number 936/21209718/A dated July 2008
SIRA Report	Report number 70142124 dated August 2017

Certificate No : Sira MC030013/11
This Certificate issued : 06 March 2018

*This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Product Certified

The measuring system consists of the following parts:



1. Sample Probe	2. Heated Filter	3. Heated Sample Line	4. Gas Conditioning	5. Analyser
Model: JCT-JES 301 / Buhler 222.20 Heated filter Probe	Model: N/A – Integrated with sample probe	Model: JCT-JH3F Length: To suit application	Model: Bespoke system with JCT / Universal Analyser coolers	Model: 4900 Analyser

Allowable variations could include:

- A different brand or model of sampling system of the same type, provided that there is evidence the alternative system works with similar types of CEM.
- Additional manifolds and heated valves used to allow more than one analyser to share a sampling system.

This certificate applies to all instruments fitted with software version 4X00/CP0_08 onwards (serial number 653043 onwards).

Certificate No : Sira MC030013/11
This Certificate issued : 06 March 2018

*This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Certified Performance

The instrument was evaluated for use under the following conditions:

Ambient Temperature Range: Stack mounted components -10°C to +55°C
Control Unit +5°C to +45°C
Instrument IP rating: IP20

Note: For outdoor installations the analyser needs to be mounted into an IP65 environment. If the instrument is supplied with an enclosure, then the ambient temperature shall be monitored inside the enclosure to ensure that it stays within the above ambient temperature range.

Results are expressed as error % of certification ranges CO 0 to 75mg/m³, NO 0 to 268mg/m³, SO₂ 0 to 572 mg/m³ and O₂ 0 to 25%vol, unless otherwise stated.

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Response time						
CO					34s	<200s
NO					35s	<200s
SO ₂					48s	<200s
O ₂					26s	<200s
Repeatability standard deviation at zero point						
CO		0.93				<2.0%
NO	0.07					<2.0%
SO ₂	0.13					<2.0%
O ₂	0.00					<0.2%
Repeatability standard deviation at reference point						
CO	0.07					<2.0%
NO	0.22					<2.0%
SO ₂	0.13					<2.0%
O ₂	0.08					<0.2%
Lack-of-fit						
CO	0.4					<2.0%
NO	0.4					<2.0%
SO ₂	0.5					<2.0%
O ₂	0.1					<0.2%

Certificate No : Sira MC030013/11
This Certificate issued : 06 March 2018

*This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Influence of ambient temperature zero point -10°C to +55°C						
CO			-1.5			<5.0%
NO		-0.8				<5.0%
SO ₂			1.6			<5.0%
O ₂	0.01					<0.50%
Influence of ambient temperature reference point -10°C to +55°C						
CO			1.31			<5.0%
NO			-1.8			<5.0%
SO ₂			-1.7			<5.0%
O ₂	-0.04					<0.50%
Influence of sample gas flow for extractive CEMS						
CO		0.66				<2.0%
NO	-0.25					<2.0%
SO ₂	0.31					<2.0%
O ₂	-0.17					<0.2%
Influence of voltage variations 190V to 250V						
CO	0.15					<2.0%
NO	0.00					<2.0%
SO ₂	0.19					<2.0%
O ₂	0.07					<0.2%

Certificate No : Sira MC030013/11
This Certificate issued : 06 March 2018

*This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Cross-sensitivity at zero with interferents: O ₂ , H ₂ O, CO, CO ₂ , CH ₄ , N ₂ O, NO, NO ₂ , NH ₃ , SO ₂ , HCl,						
CO			1.80			<4.0%
NO				3.30		<4.0%
SO ₂			1.30			<4.0%
O ₂	0.00					<0.40%
Cross-sensitivity at reference with interferents: O ₂ , H ₂ O, CO, CO ₂ , CH ₄ , N ₂ O, NO, NO ₂ , NH ₃ , SO ₂ , HCl,						
CO			1.80			<4.0%
NO		0.60				<4.0%
SO ₂				2.70		<4.0%
O ₂	0.00					<0.40%
Measurement uncertainty					Guidance - at least 25% below max permissible uncertainty	
CO					5.5%	<7.5% (10%)
NO					7.3%	<15% (20%)
SO ₂					13.7%	<15% (20%)
O ₂					2.3%	<30% (40%)
Calibration function (field)						
CO					>0.999	>0.90
NO					>0.999	>0.90
SO ₂					>0.999	>0.90
O ₂					>0.999	>0.90
Response time (field)					Note1	
CO					34s	<200s
NO					35s	<200s
SO ₂					48s	<200s
O ₂					26s	<200s

Certificate No : Sira MC030013/11
This Certificate issued : 06 March 2018

*This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Lack of fit (field)					Note 2	
CO			<2			<2.0%
NO			<2			<2.0%
SO ₂			<2			<2.0%
O ₂			<2			<0.2%
Maintenance interval	With SO ₂ = 2 weeks Without SO ₂ = 3 weeks					>8 days
Zero and Span drift requirement	Compensation for zero and span drift is performed by introducing suitable zero and span gases as described in the 4900 analyser manual					Clause 6.13 & 10.13 Manufacturer shall provide a description of the technique to determine and compensate for zero and span drift.
Change in zero point over maintenance interval						
CO			<2.0		2 weeks	<3.0%
NO			1.2		2 weeks	<3.0%
SO ₂				2.1	3 weeks	<3.0%
O ₂				2.8	2 weeks	<0.2%
Change in reference point over maintenance interval						
CO						<3.0%
NO	0.4				2 weeks	<3.0%
SO ₂	0.4				3 weeks	<3.0%
O ₂			1.2		2 weeks	<0.2%
Availability					98.9%	>95% (>98% for O ₂)

Certificate No : Sira MC030013/11
This Certificate issued : 06 March 2018

*This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Test	Results expressed as % of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<5		
Reproducibility						
CO			1.34			<3.3%
NO		1.0				<3.3%
SO ₂			1.32			<3.3%
O ₂	0.41					<0.20%

Note 1. Response time stated is from the lab test.

Note 2. Data derived from the calibration function test (field)

Certificate No : Sira MC030013/11
This Certificate issued : 06 March 2018

*This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*

Description

The Servomex 4900 samples flue gas via an extractive process. The analyser measures oxygen by a magnetodynamic paramagnetic sensor and NO, CO and SO₂ by infrared gas filter correlation technology.

The Servomex 4900 series system tested consisted of a Servomex 4995 Sample Conditioning System and two Servomex 4900 analysers to both measure NO, CO, SO₂ and O₂. The system also included a heated filter probe model JES 300 (located within the stack) and a heated line model JH3F, both are manufactured by JCT Consulting & Trading GmbH. The configuration of analysers and sample conditioning system can be varied depending upon the application and customer requirements.

General Notes

1. This certificate is based upon the equipment tested. The Manufacturer is responsible for ensuring that on-going production complies with the standard(s) and performance criteria defined in this Certificate. The Manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management system shall be subject to regular surveillance according to 'Regulations Applicable to the Holders of Sira Certificates'. The design of the product certified is defined in the Sira Design Schedule V06 for certificate No. Sira MC130013/11.
2. If certified product is found not to comply, Sira Certification Service should be notified immediately at the address shown on this certificate.
3. The Certification Marks that can be applied to the product or used in publicity material are defined in 'Regulations Applicable to the Holders of Sira Certificates'.
4. This document remains the property of Sira and shall be returned when requested by the company.

Certificate No : Sira MC030013/11
This Certificate issued : 06 March 2018

*This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts*