

MCERTS Certified Products: Continuous Water Monitoring System (CWMS) Part 3 – Water Flowmeters



Expired Certificates

The flowmeters listed below are no longer MCERTS certified. The renewal date on the certificate indicates the point at which the certification lapsed. Any product manufactured after this date is not certified.

For instruments already installed, in use or in the supply chain before the renewal date on the products' MCERTS certificate are not affected by certificate expiry.

Products on this list may be undergoing recertification, please contact the manufacturer directly for more information.

For current MCERTS certified flowmeters, please visit:

<http://www.siraenvironmental.com/UserDocs/mcerts/MCERTSCertifiedProductsCWMSPart3.pdf>

Certificate holder	Model	Certified range	Certificate No.
GE Sensing EMEA Tel: +353 61 470200 www.gesensing.com GE Sensing (UK agent) Tel: +44 (0)116 231 7100	Panametrics AT868 Ultrasonic Liquid Flowmeter	0.25 to 5 m/s	Sira MC120215/00
Nivus GmbH www.nivus.com	OCM Pro CF (incorporating POA Sensor) Echo Profile Flowmeter	Velocity 0 to 2 m/s Depth 0.1 to 1 m	Sira MC110189/02



MCERTS Certified Products: Continuous Water Monitoring System (CWMS) Part 3 – Water Flowmeters



Expired Certificates

Certificate holder	Model	Certified range	Certificate No.
Nivus GmbH www.nivus.com	OCM Pro CF (incorporating POA pipe sensor) Ultrasonic Echo Profile Flowmeter	Velocity 0.05 to 2.5 m/s Depth 0.075 to 0.6 m	Sira MC130219/01
Siemens Milltronics Process Instruments Inc. Siemens Automation & Drives Tel: +44 (0)8458 507600 www.siemens.com/processinstrumentation	OCM III non-contact level flow monitor with TS-2 Temperature Sensor & Echomax XRS-5 ultrasonic transducer	0 to 2 metres (nominal)	Sira MC050058/04

