

MCERTS Certified Products: Continuous Water Monitoring System (CWMS) Part 2 – On-line analysers



Any potential user should ensure, in consultation with the manufacturer, that the CWMS is suitable for the process on which it will be installed. It is important to refer to the details on certificates issued as they may contain specific limitations on use. The certificate holders listed have submitted satisfactory test reports and successfully undergone a product manufacturing audit.

Certificate holder	Model	Certified range	Certificate No.
ABB Ltd Tel: +44 (0)1453 826661 www.abb.com	ABB AX 400 pH Measurement System	pH units 0 to 12.5	Sira MC120217/03 Version issued: 10/01/2018
ABB Ltd Tel: +44 (0)1453 826661 www.abb.com	ABB ATS430 Turbidity Probe with AWT400 Controller	Turbidity 0 to 50 NTU 0 to 500 NTU 0 to 4000 NTU	Sira MC160305/00 Version issued: 19/08/2016
BioTector Analytical Systems Ltd Tel: +353 (0) 21 4374237 Email: info@biotector.com Website: www.biotector.com	Hach BioTector B7000 TOC Analyzer	TOC 0 to 50 mg/L 0 to 1000 mg/L 0 to 10000 mg/L	Sira MC120199/04 Version issued: 08/11/2017
BioTector Analytical Systems Ltd Tel: +353 (0) 21 4374237 Email: info@biotector.com Website: www.biotector.com	Hach BioTector B3500 TOC Analyzer	TOC 0 to 25 mg/L	Sira MC130228/02 Version issued: 24/07/2018



MCERTS Certified Products: Continuous Water Monitoring System (CWMS) Part 2 – On-line analysers



Any potential user should ensure, in consultation with the manufacturer, that the CWMS is suitable for the process on which it will be installed. It is important to refer to the details on certificates issued as they may contain specific limitations on use. The certificate holders listed have submitted satisfactory test reports and successfully undergone a product manufacturing audit.

Certificate holder	Model	Certified range	Certificate No.
Endress & Hauser Conducta GmbH UK Office: Tel: +44(0)161 286 5000 Fax: +44(0)161 286 5098 info@uk.endress.com www.uk.endress.com	CM44* transmitter with CPS11D pH Sensor	pH units 2 to 10	Sira MC140246/01 Version issued: 09/04/2014
Hach Lange GmbH Tel: 0161 872 1487 Email: sales@hach-lange.co.uk www.hach-lange.co.uk	SC100 controller unit SC1000 controller unit Probes: SOLITAX sc, pHD sc, LDO	pH 1 to 13 DO 0 to 200%sat 0 to 20 mg/L Turbidity 0 to 1000 NTU	Sira MC120214/01 Version issued: 15/11/2017
Pollution & Process Monitoring Ltd Tel: +44(0)1732 882044 www.pollution-ppm.co.uk	Protoc 300 TOC Analyser	TOC 0 to 20mg/L 0 to 4000 mg/L	Sira MC060077/05 Version issued: 11/08/2016
Shimadzu Corporation Tel: +44(0)1908 552209 Email: info@shimadzu.co.uk www.shimadzu.co.uk	Shimadzu On-line TOC Analyzer TOC-4200	TOC 0 to 100 mg/L 0 to 500 mg/L 0 to 1000 mg/L 0 to 10,000 mg/L	Sira MC160311/01 Version issued: 22/02/2017



MCERTS Certified Products: Continuous Water Monitoring System (CWMS) Part 2 – On-line analysers



Any potential user should ensure, in consultation with the manufacturer, that the CWMS is suitable for the process on which it will be installed. It is important to refer to the details on certificates issued as they may contain specific limitations on use. The certificate holders listed have submitted satisfactory test reports and successfully undergone a product manufacturing audit.

Certificate holder	Model	Certified range	Certificate No.
Xylem Analytics Germany GmbH Tel: +44(0)1462 673581 www.xylemanalytics.co.uk	IQ SENSOR NET 181, Controller Unit FDO 700 IQ Optical Dissolved Oxygen Probe / VisoTurb 700 IQ Turbidity Probe	DO 0 to 20mg/L 0 to 200% saturation Turbidity 0 to 100 NTU 0 to 500 NTU	Sira MC150286/01 Version issued: 01/07/2016
Xylem Analytics Germany GmbH Tel: +44(0)1462 673581 www.xylemanalytics.co.uk	IQ SENSOR NET 182, 182 XT & 182 XT-4 Controller Unit FDO 700 IQ Optical Dissolved Oxygen Probe / VisoTurb 700 IQ Turbidity Probe	DO 0 to 20mg/L 0 to 200% saturation Turbidity 0 to 100 NTU 0 to 500 NTU	Sira MC150280/01 Version issued: 01/07/2016
Xylem Analytics Germany GmbH Tel: +44(0)1462 673581 www.xylemanalytics.co.uk	IQ SENSOR NET 2020XT, 2020XT-H3 & 2020XT-H3-C6 Controller Unit FDO 700 IQ Optical Dissolved Oxygen Probe / VisoTurb 700 IQ Turbidity Probe	DO 0 to 20mg/L 0 to 200% saturation Turbidity 0 to 100 NTU 0 to 500 NTU	Sira MC150281/01 Version issued: 01/07/2016

