

MCERTS Certified Products: Indicative Ambient Particulate Monitors



Any potential user should ensure, in consultation with the manufacturer, that the CAMS is suitable for the application. It is important to refer to the details on certificates issued as they may contain specific limitations on use. The certificate holders listed have submitted satisfactory test reports and successfully undergone a product manufacturing audit.

| Certificate holder | Model | Certified range | Certificate No. |
|--|--|---|--|
| Aeroqual Ltd Tel: +64 9 623 3013 www.aeroqual.com | Dust Sentry | PM ₁₀ 0 to 150µg/m ³ | Sira MC 130235/01 Version issued: 19/10/2018 |
| Aeroqual Ltd Tel: +64 9 623 3013 www.aeroqual.com | AQM 65 | PM ₁₀ 0 to 150µg/m ³ | Sira MC 160289/00 Version issued: 16/03/2016 |
| MetOne Instruments Inc www.metone.com UK Sales (EnviroTechnology Services plc): Tel: +44 (0) 1453 733200 www.et.co.uk info@et.co.uk | E-Sampler Real Time Aerosol Monitor ES-642 Real Time Aerosol Monitor | PM _{2.5} 0 to 150µg/m ³ PM ₁₀ 0 to 150µg/m ³ | Sira MC 130241/03 Version issued: 14/12/2018 |
| TSI Incorporated UK Sales: Tel: +44 (0) 149 4 459200 www.tsi.com | Environmental DustTrak™ Aerosol Monitor EDTPM10M | PM ₁₀ 0 to 10,000 µg/m ³ | Sira MC160316/02 Version issued: 19/12/2018 |
| TSI Incorporated UK Sales: Tel: +44 (0) 149 4 459200 www.tsi.com | Environmental DustTrak™ Aerosol Monitor EDTPM2.5M | PM _{2.5} 0 to 10,000 µg/m ³ | Sira MC160317/02 Version issued: 19/12/2018 |

MCERTS Certified Products: Indicative Ambient Particulate Monitors



| Certificate holder | Model | Certified range | Certificate No. |
|--|---|---|--|
| TSI Incorporated UK Sales: Tel: +44 (0) 149 4 459200 www.tsi.com | Environmental DustTrak™ Aerosol Monitor EDTDRXM | PM _{2.5} 0 to 10,000 µg/m ³ PM ₁₀ 0 to 10,000 µg/m ³ | Sira MC160318/02 Version issued: 19/12/2018 |
| Turnkey Instruments Ltd Tel: +44(0)1606 330020 www.turnkey-instruments.com | Osiris Airborne Particulate Monitor | PM ₁₀ 0 to 100µg/m ³ | Sira MC 090157/05 Version issued: 21/08/2015 |
| Turnkey Instruments Ltd Tel: +44(0)1606 330020 www.turnkey-instruments.com | Topas Airborne Particulate Monitor | PM ₁₀ 0 to 100µg/m ³ | Sira MC 090158/05 Version issued: 21/08/2015 |