

MCERTS Certified Products: Portable Water Monitoring System (PWMS)



Any potential user should ensure, in consultation with the manufacturer, that the PWMS is suitable for the process on which it will be installed. It is important to refer to the details on certificates issued as they may contain specific limitations on use. The certificate holders listed have submitted satisfactory test reports and successfully undergone a product manufacturing audit.

Certificate holder	Model	Certified range	Certificate No.
Hach Lange GmbH Tel: 0161 872 1487 Email: sales@hach-lange.co.uk www.hach-lange.co.uk	HQ40D portable meter Probes: pH and conductivity	pH: 1 to 13 Conductivity: 0.01 to 500µS/cm Conductivity: 0.01 to 200mS/cm	Sira MC120213/01 Version issued: 15/11/2017

