

MCERTS Certified Products: Portable Water Monitoring System (PWMS)



Expired Certificates

The portable water monitoring systems listed below are no longer MCERTS certified. The renewal date on the certificate indicates the point at which the certification lapsed. Any product manufactured after this date is not certified.

For instruments already installed, in use or in the supply chain before the renewal date on the products' MCERTS certificate are not affected by certificate expiry.

Products on this list may be undergoing recertification, please contact the manufacturer directly for more information.

For current MCERTS certified portable water monitoring systems, please visit
www.siraenvironmental.com/UserDocs/mcerts/MCERTSCertifiedProductsPWMS.pdf

Certificate holder	Model	Certified range	Certificate No.
Hach UK Supplier: OTT Hydrometry Tel: +44 (0)1246 573 480 Email: Sales@ott-hydrometry.co.uk www.ott-hydrometry.co.uk	Hydrolab DS5, DS5X & MS5 Sonde with Field Rugged PDA unit	pH: 2 to 12 Temperature: 0 to 45°C Turbidity: 0 to 500 NTU Conductivity: 0 to 1 mS/cm Conductivity: 0 to 100 mS/cm	Sira MC100181/00 MCERTS certification only valid for products manufactured before 08/11/2015
YSI Inc. Tel: +44(0)1462 673 581 www.yasihydrodata.com	YSI Professional Plus Portable Multiparameter Water Quality Meter	DO (polarographic) : 0 to 200% saturation pH: 2 to 12 Conductivity: 0 to 1 mS/cm Conductivity: 0 to 50 mS/cm	Sira MC100179/01 MCERTS certification only valid for products manufactured before 12/03/2019

