





# PRODUCT CONFORMITY CERTIFICATE

This is to certify that the

# **DUSTHUNTER T200**

Manufactured by:

# SICK AG

Bergener Ring 27 01458 Ottendorf-Okrilla Germany

has been assessed by Sira Certification Service And for the conditions stated on this certificate complies with:

MCERTS Performance Standards for Continuous Emission Monitoring Systems (CEMS), Version 4 dated July 2018 EN15267-1:2009, EN15267-2:2009, EN15267-3:2007,

& QAL 1 as defined in EN 14181: 2014

## Certification Ranges:

Dust 0 to 0.1 Ext.\*

0 to 0.05 Ext. 0 to 0.2 Ext. 0 to 0.5 Ext. 0 to 1.0 Ext.

**Emily Alexander** 

\*0 to 0.1 Ext. ≡ 0-15 mg/m³ dust at 5m optical path length

Project No. : 674/0391F /70219144
Certificate No : Sira MC090145/03
Initial Certification : 24 April 2009
This Certificate issued : 23 April 2019

Renewal Date : 23 April 2024 Environmental Project Engineer

MCERTS is operated on behalf of the Environment Agency by

## **Sira Certification Service**



Unit 6, Hawarden Industrial Park Hawarden, Deeside, CH5 3US Tel: +44 (0)1244 670 900

The MCERTS certificate consists of this document in its entirety.
For conditions of use, please consider all the information within.
This certificate may only be reproduced in its entirety and without change
To authenticate the validity of this certificate please visit www.csagroupuk.org/mcerts







#### **Certificate Contents**

Approved Site Application	2
Basis of Certification	
Product Certified	
Certified Performance	
Description	5
General Notes	

## **Approved Site Application**

Any potential user should ensure, in consultation with the manufacturer, that the monitoring system is suitable for the intended application. For general guidance on monitoring techniques refer to the Environment Agency Monitoring Technical Guidance Notes available at <a href="https://www.mcerts.net">www.mcerts.net</a>

On the basis of the assessment and the ranges required for compliance with EU Directives this instrument is considered suitable for use on waste incineration and large coal-fired combustion plant applications. This CEM has been proven suitable for its measuring task (parameter and composition of the flue gas) by use of the QAL 1 procedure specified in EN14181, for LCPD and WID applications for the ranges specified. The lowest certified range for each determinand shall not be more than 1.5X the daily average emission limit value (ELV) for WID applications, and not more than 2.5X the ELV for LCPD and other types of application.

The field trial was conducted over 15 months on a municipal waste incinerator.

#### **Basis of Certification**

This certification is based on the following Test Report(s) and on Sira's assessment and ongoing surveillance of the product and the manufacturing process:

TÜV Rhineland Report Number 936/2120461/F dated 18.03.2008

#### **Product Certified**

The Dusthunter T200 measuring system consists of the following parts:

- Sender/receiver (SR) unit DHT-T21
- Connection cable to connect SR unit to the control unit
- Reflector DHT-R1x
- Connection cable to connect the reflector to the SR unit
- Control unit MCU for data control, evaluation and output
  - With integrated purge air supply, for internal duct pressure -50... +2 mbar
  - Without purge air unit, therefore additionally required:
- Optional external purge air unit, for internal duct pressure -50... 30 mbar

This certificate applies to all instruments fitted with software version 1.026 (MCU) 1.3.04 (sensor) and 02.16 (SOPAS ET operating software), serial number 07478656 (SR unit) 07478637 (MCU) 07478660 (R/SL measuring device) onwards.

Certificate No: Sira MC090145/03 This Certificate issued: 23 April 2019







## **Certified Performance**

The instrument was evaluated for use under the following conditions:

Ambient Temperature Range: -20°C to +50°C

Instrument IP rating: IP66

Note: If the instrument is supplied with an enclosure, then the ambient temperature shall be monitored inside the enclosure to ensure that it stays within the above ambient temperature range.

Unless otherwise stated the evaluation was carried out on the certification range 0 to 0.1 Ext.

Test	Test Results expressed as % of certification range				Other results	MCERTS specification	
	<0.5	<1	<2	<5		•	
Response time					28s (with integration time set to 30s)	<200s	
Repeatability standard deviation at zero point	0.1					<2.0%	
Repeatability standard deviation at reference point	0.1					<5.0%	
Lack-of-fit							
0 to 0.1 Ext.			1.0			<3.0%	
0 to 0.2 Ext.		0.8				<3.0%	
0 to 0.5 Ext.		0.7				<3.0%	
0 to 1.0 Ext.	0.1					<3.0%	
Influence of ambient temperature zero point (-20°C to +50°C)		-0.9				<5.0%	
Influence of ambient temperature reference point (-20°C to +50°C)		0.7				<5.0%	
Influence of voltage variations (190V to 250V)		0.7				<2.0%	
Influence of vibration (10 to 60Hz (±0.3mm), 60 to 150Hz at 19.6m/s²)	0.3					To be reported	
Measurement uncertainty					Guidance - at least 25% below max permissible uncertainty		
(For and ELV of 10 mg/m <sup>3</sup> )					8.4%	<<22.5% (30%)	
Calibration function (field)					0.89 Note 1	>0.90	

Certificate No : Sira MC090145/03 This Certificate issued : 23 April 2019







Test Results expressed as % of the certification range					Other results	MCERTS specification
	<0.5	<1	<2	<5		-
Response time (field)					28s (with integration time set to 30s)	<200s
Lack of fit (field)			1.8			<3.0%
Maintenance interval					6 months Note 2	>8 days
Zero and Span drift requirement	Zero V The set so that drifts of the over the 'zer  Control Sender % during the istandate error is error in cycle determ evaluate control swivell	Clause 6.13 & 10.13  Manufacturer shall provide a description of the technique to determine and compensate for zero and span drift.				
Change in zero point over maintenance interval		-0.5				<3.0%
Change in reference point over maintenance interval				-2.9		<3.0%
Availability					99.3%	>95%
Reproducibility				2.9		<3.3%

Note 1: The calibration function result / R2 values are between 0.8 and 0.9 due to low dust levels. The CEMS pass the EN14181 criteria, but not the requirement for EN15267-3.

Note 2: The T200 has a maintenance interval of 6 months. In the case of a new installation the measuring system should be tested by all means in weekly or biweekly intervals via visual inspection.

The work detailed below has to be carried out at regular intervals, depending on local conditions:

- Visual inspection of the CEM
- Examination of the S/R unit and the reflector by swing out and visual inspection. The optical surfaces should be cleaned
  if necessary.
- Determination of zero and span point
- Examination of the purge air supply
- Check cycle operation including a check of zero and span point and of the contamination signal.

Certificate No : Sira MC090145/03 This Certificate issued : 23 April 2019







### **Description**

The DUSTHUNTER T200 uses transmission measurement to determine the mass concentration of dust in flowing gases. The measuring system operates as a transmitter with a double beam path, with two-sided detection by a sender/receiver (SR) unit and a reflector.

A high performance LED sends light in the visible range through the active measuring path containing particles to the reflector where it is bounced back to the receiver. While passing through the measurement path twice, the transmitted light is attenuated by the particles within the path and then captured by the measurement receiver. Continuous monitoring of the sender output registers the smallest changes in brightness of the transmitted light beam which serves to determine the measurement signal.

#### **General Notes**

- 1. This certificate is based upon the equipment tested. The Manufacturer is responsible for ensuring that on-going production complies with the standard(s) and performance criteria defined in this Certificate. The Manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management system shall be subject to regular surveillance according to 'Regulations Applicable to the Holders of Sira Certificates'. The design of the product certified is defined in the Sira Design Schedule for certificate No. Sira MC090145/00
- 2. If certified product is found not to comply, Sira Certification Service should be notified immediately at the address shown on this certificate.
- 3. The Certification Marks that can be applied to the product or used in publicity material are defined in 'Regulations Applicable to the Holders of Sira Certificates'.
- 4. This document remains the property of Sira and shall be returned when requested by the company.

Certificate No : Sira MC090145/03 This Certificate issued : 23 April 2019